

भारत सरकार /Government of India

स्वास्थ्य एवं परिवार कल्याण मंत्रालय / Ministry of Health and Family Welfare

स्वास्थ्य सेवा महानिदेशालय /Directorate General of Health Services

**आखिल भारतीय भौतिक चिकित्सा एवं पुनर्वास संस्थान**

**ALL INDIA INSTITUTE OF PHYSICAL MEDICINE AND REHABILITATION**

हाजी अली,केमार्ग .के., महालक्ष्मी,मुंबई 400034

Haji Ali, K.K.Marg, Mahalaxmi, Mumbai – 400 034.

संदर्भ सं./ Ref.No.: Estt./Quot/OT/334

दिनांक / Dated : 12/11/ 2021

To,

महोदय/Sir,

**Tubing for Therapeutic Exercise, Pediatric Pinch Gauge, Ball Pool, Bolsters of Various Sizes, Thera Band, Flexi bar, Thera Putty, Pediatric Hand Dynamometer** की खरीदी निम्न प्रकार से प्रस्तावित की जाती है।

It is proposed to purchase **Tubing for Therapeutic Exercise, Pediatric Pinch Gauge, Ball Pool, Bolsters of Various Sizes, Thera Band, Flexi bar, Thera Putty, Pediatric Hand Dynamometer** as mentioned below.

उपर्युक्त आपूर्ति के लिए जीएसटी सहित न्यूनतम दर प्रस्तुत करें।

Submit lowest rate for supply of the same, including GST

दरें बताते समय कृपया निम्नलिखित जानकारी दें :

While quoting, the following information may please be supplied:

- वस्तुओं का पूर्ण विवरण-वस्तुओं को सचित्र स्पष्ट करनेवाली पुस्तिकाओं, इशतेहारों या नमूनों के रूप में।  
Complete description of the articles in form of booklets, pamphlets or samples illustrating the articles.
- अवधि, जिसमें आपूर्ति किया जा सके।  
The period within which the supply can be made.
- जीएसटी संख्या/कर इत्यादि, यदि लागू हों।  
GST number/taxes etc. if applicable.

उपरोक्त निविदा को मुहरबंद लिफाफे में भरकर, उसपर शिर्षक **'Tubing for Therapeutic Exercise, Pediatric Pinch Gauge, Ball Pool, Bolsters of Various Sizes, Thera Band, Flexi bar, Thera Putty, Pediatric Hand Dynamometer के लिये निविदा'** लिखकर, उसे "निदेशक, आखिल भारतीय भौतिक चिकित्सा एवं पुनर्वास संस्थान, हाजी अली, मुंबई-400 034" को डाक द्वारा निर्धारित समय के अंदर भेजें या स्थापना अनुभाग में उपलब्ध निविदा पेट्टी में दिनांक **29/11/2021** में 12.30 बजे (अपराहन) के पूर्व डालें।

The above mentioned quotation should be kept in a sealed envelope, superscribed as "**Quotation for Tubing for Therapeutic Exercise, Pediatric Pinch Gauge, Ball Pool, Bolsters of Various Sizes, Thera Band, Flexi bar, Thera Putty, Pediatric Hand Dynamometer**" & should be sent by post within the prescribed time to the "**Director, All India Institute of Physical Medicine and Rehabilitation, Haji Ali Park, Mumbai – 400 034**" or may be dropped in quotation Box available with Establishment Section before 12.30 P.M. (I.S.T.) on **29/11/2021**.

निविदा, तभी वैध मानी जाएगी जब उसमें प्राधिकृत हस्ताक्षरकर्ता के नाम एवं पदनाम, एवं उनके पैन/जीएसटी संख्या सही लिखी गई हो।

Quotation will only be valid when name & designation of authorized signatory and his PAN/GST Number are mentioned correctly.

निविदा एक वर्ष की अवधि के लिए वैध होनी चाहिए।

Quotations should be valid for the period of one year.

इस संबंध में किसी प्रकार की अन्य जानकारी आवश्यक हो तो आप इस कार्यालय में सोमवार से शुक्रवार, 9.00 (पूर्वाहन) से 3.30 (अपराहन) बजे के बीच में प्राप्त कर सकते हैं।

Any other information you may require in this matter can be obtained from this office from Monday to Friday between 9.00 a.m. to 3.30 p.m.

Particular
List Attached

संस्थान की जीएसटी संख्या / GST No. of the Institute : (27MUMA28322A1D1)

Sd/-

(मुख्य चिकित्सा अधिकारी (एन.एफ.एस.जी)  
(Chief Medical Officer (NFSG))

सूचना / Note :

- पूर्व रसीदी बिल के तीन प्रतियों के अनुसार भुगतान, किया जाएगा।  
Payment will be made against pre-receipted bill in triplicate.
- खरीदी जाने वाली वस्तु GeM प्लेटफॉर्म पर उपलब्ध नहीं है।  
Item to be purchased is not available on GeM Platform.

### Tubing for Therapeutic Exercise

0

<u>Sr. No.</u>	<u>Particular</u>		<u>QTY</u>
	<u>Tubing for Therapeutic Exercise</u>		
1	Red	Light	01-20ft
2	Green	Medium	01-20ft
3	Blue	Heavy	01-20ft
4	Black	x-Heavy	01-20ft
	<b>Specification :-</b> 1. Material should be clean, non-toxic and bleed proof. 2. Material should not fragment, separate or stick to the skin. 3. Should be easy to carry and convenient for storing. 4. Should be easily identifiable in strength by its colour. 5. Should have long shelf-life (at least 4-5 yrs.)		

### Pediatric Pinch Gauge

<u>Sr. No.</u>	<u>Particular</u>	<u>QTY</u>
	<u>Pediatric Pinch Gauge</u>	
	<b>Specification: The Pinch Gauge should</b>	
1	Display pinch force in pounds and kilograms.	02
2	Have pinch grasp which is comfortable and effective.	
3	Have readings which are displayed accurately	
4	Have standardized calibration	
5	Be smooth to handle and should be durable.	

### Ball Pool

<u>Sr. No.</u>	<u>Particular</u>	<u>QTY</u>
	<u>Ball Pool</u>	
	<b>Specification:</b>	
1	Should be expandable, portable & safe.	01
2	Should have design which allows for shape flexibility	
3	Panels (48" L x 26" H) are solid with foam padding & reinforced vinyl cover.	
4	Sensory environment pool gives child visual & tactile feedback, provides the buoyancy of water without the mess.	
5	Required – 6 panels of 5 x 7 Feet with required balls 5500.	
6	It should be colorful.	

### Bolsters of Various Sizes

<u>Sr. No.</u>	<u>Particular</u>		<u>QTY</u>
	<u>Bolsters of Various Sizes</u>		
	<b>Specification:</b>		
	Should have soft yet firm rolls (Bolsters)		
	Have a solid core to prevent flattening at the bottom.		
	Adaptive shapes for rolling & rocking movements for positioning & neurodevelopmental training.		
	<b>Length</b>	<b>Diameter</b>	01 for each size
1	24"	4"	
2	24"	6"	
3	24"	8"	
4	36"	10"	

### Thera Band

<u>Sr. No.</u>	<u>Particular</u>		<u>QTY</u>
	<u>Thera Band</u>		
1	Yellow	Light	46 mtr. Length - One
2	Red	Medium	46 mtr. Length - One
3	Blue	x-Heavy	46 mtr. Length - One
4	Black	xx-Heavy	46 mtr. Length - One
	<b>Specification :-</b> 1. Theraband material should be latex free. 2. Resistance levels of the band should be color-coded 3. Bands should have good tensile strength, elasticity and should be durable.		

**Flexi Bar**

<u>Sr. No.</u>	<u>Particular</u>			<u>QTY</u>
	<u>Flexi Bar</u>			
1	Tan	xx-Light	12" Long	01
2	Yellow	x- Light	12" Long	01
3	Red	Light	12" Long	01
4	Green	Medium	12" Long	01
5	Blue	Heavy	12" Long	01
6	Black	x-Heavy	12" Long	01
<b><u>Specification :-</u></b> 1. Material should be clean, non-toxic and safe to use in various neurological and musculoskeletal conditions. 2. Should be lightweight & portable easy to store. 3. Should not fragment, separate or stick to the skin. 4. Should have long shelf-life (at least 4-5 yrs.) 5. Should be easily identifiable in strength by its colour.				

**Thera Putty**

<u>Sr. No.</u>	<u>Particular</u>		<u>QTY</u>
	<u>Thera Putty</u>		
1	Therapy Putty	Soft yellow	One of 454 GMS
2	Therapy Putty	Medium Soft Red	One of 454 GMS
3	Therapy Putty	Medium Green	One of 454 GMS
4	Therapy Putty	Firm Blue	One of 454 GMS
5	Therapy Putty	Extra Firm Grey	One of 454 GMS
<b><u>Specification:</u></b> 1. Material should be clean, non-toxic, non-oily and Bleed-proof. 2. Material should not fragment, separate or stick to the skin. 3. Material should meet wide range of strengthening needs. 4. Should have long shelf-life (at least 4-5 yrs.)			

**Pediatric Hand Dynamometer**

<u>Sr. No.</u>	<u>Particular</u>	<u>QTY</u>
	<u>Pediatric Hand Dynamometer</u>	
<b><u>Specification: The Dynamometer should</u></b>		
1	Be able to measure the maximum isometric strength of the hand and forearm muscles.	02
2	Display grip force in pounds and kilograms.	
3	Have Grasp handle which is comfortable to hold.	
4	Have readings which are displayed accurately.	
5	Have handle to accommodate various sizes of hands and 3-5 hand adjustable Placements.	
6	Be useful for measuring Right and Left hand.	
7	Be calibratable regularly for Consistent results.	
8	Minimum warranty for 2 years should be given.	

Sd/-

(मुख्य चिकित्सा अधिकारी (एन.एफ.एस.जी)  
 (Chief Medical Officer (NFSG))